



OPERATING INSTRUCTION MANUAL

AS SERIES CONDUCTIVITY CELLS

REV. 4

Water Analytics, Inc.
100 School Street
Andover, MA 01810

978-749-9949 Toll free: 855-747-7623
Fax: 978-749-9961
www.WaterAnalytics.net

TABLE OF CONTENTS

1.0 GENERAL INFORMATION	3
1.1 Introduction.....	3
1.2 Cautions To Be Observed	3
2.0 SPECIFICATIONS.....	4
3.0 PRINCIPLE OF OPERATION	5
4.0 INSTALLATION	6
4.1 Installation of Flow-Through Sensors.....	6
4.2 Installation of Submersion Sensor	6
6.0 TROUBLESHOOTING AND SERVICE	8
7.0 DRAWINGS.....	9
7.1 N106-280 AS SERIES CONTACTING CONDUCTIVITY CELLS	9
7.2 N105-60 INSTALLATION OF AS SERIES CONTACTING CONDUCTIVITY CELLS.....	10
7.3 N105-61 ASSEMBLY OF AS SERIES CONTACTING CONDUCTIVITY CELLS COMPRESSION FITTING.....	11

AS SERIES CONDUCTIVITY CELLS INSTRUCTION MANUAL

1.0 GENERAL INFORMATION

1.1 Introduction

- 1.1.1 This manual provides information on AquaMetrix Model AS Series flow-through and submersion sensors. Conductivity sensors have long been called cells. The two terms, cell and sensor, are used interchangeably in this manual.

1.2 Cautions To Be Observed

- 1.2.1 Consult the factory before using sensor in extremely strong solvents such as ethylene dichloride. Also, do not exceed the sensor's maximum ratings listed in Section 2 of this manual.
- 1.2.2 Before placing sensor into service, remove protective plastic cap.
- 1.2.3 Do not exceed the maximum sensor-to-analyzer distance specified for the analyzer you are using.
- 1.2.4 Be sure the cell constant is appropriate for the scaling of the analyzer. The constant is given in the numerical part of the model number. For example, ASM-001-xxx has a constant of 0.01; ASM-1-xxx has a constant of 1.

2.0 SPECIFICATIONS

CELL CONSTANTS: 0.01 to 50

THREADED SIZE: 3/4" NPT

TEMPERATURE COMPENSATION: Automatic 0-120°C (32-248°F)

CABLE LENGTH: 6m (20ft)

INSERTION LENGTH: Variable (see Section 4.1.5)

WETTED MATERIALS:

AS-S: 316 SS, Noryl, EPDM

AS-G: Graphite, Noryl, EPDM

SAFETY FEATURE: Shoulder on cells provides an integral safety stop

TEMPERATURE and PRESSURE:

<u>Type</u>	<u>Max. Press. (psi)</u>	<u>Max. Temp. (°C)</u>
P	50	60
	100	50
	150	20
S	250	120

3.0 PRINCIPLE OF OPERATION

- 3.1** All solutions containing water conduct electricity to some extent. The ability of the solution to conduct electricity is called “conductance” (the reciprocal of resistance). Addition of electrolytes such as salts, acids, or bases to pure water will increase the ability of the liquid to conduct electricity and hence, increase the solution’s conductance (decreases the resistance).
- 3.2** An electrolytic conductivity system measures solution conductance by using an analyzer interconnected with cable to a sensor immersed in the solution. The conductivity sensor is composed of two or more electrodes and a temperature sensor. The electrodes are in contact with the solution. The analyzer circuitry impresses an alternating voltage between the electrodes and the magnitude of the resulting current is linearly related to the solution conductivity.
- 3.3** When the temperature of a solution changes, its conductivity changes. The temperature change is compensated with a thermistor (temperature sensitive resistor) in the conductivity sensor to alter the gain of the measuring circuit. This is accomplished such that, regardless of actual solution temperature, the instrument display indicates what the solution conductivity would be if the solution temperature was 25°C (an internationally accepted reference).

4.0 INSTALLATION

4.1 Installation of Flow-Through Sensors Refer to Dwg# N105-60

- 4.1.1 Flow-through sensors are secured with the supplied compression fitting. The insertion length is adjustable within the limits shown in DWG# N106-280. If your sensor has a cross hole it must be inserted far enough to be sure the cross hole is in the solution. Use thread sealant such as Teflon tape on the sensor and the mounting hardware threads.
- 4.1.2 Install a 3/4" NPT tee into the process line. Tee material should be compatible with sensor materials of construction for best chemical resistance and pressure/temperature ratings. A 1" NPT tee with a 1x3/4" reducer may also be used.
- 4.1.3 Electrically connect sensor to the instrument as described in Section 4.3.
- 4.1.4 Remove the protective cap from the sensor and calibrate the system as described in the instrument operating instruction manual before installing sensor into the mounting tee.
- 4.1.5 Screw sensor and fitting into the tee. Locate end of the sensor or cross hole in sensor at the approximate center of the stream.

If you disassembled the cell be sure to re-assemble the components as shown in DWG# N105-61

4.2 Installation of Submersion Sensor Refer to Dwg# N105-60

- 4.2.1 Submersion sensors are designed for use in tanks, suspended from the top of the tank. Mounting hardware, model MHMSC is available from Water Analytics. It consists of 3/4" CPVC pipe, 24" long with a cable clamp and a wire-mounting bracket.
- 4.2.2 Run the sensor cable through the desired length of 3/4" CPVC pipe and screw on the cell using a thread sealant to keep the cable end of the sensor dry.
- 4.2.3 Electrically connect the sensor to the instrument as described in Section 4.3.
- 4.2.4 Remove the protective cap from the sensor and calibrate the system as described in the instrument's operating instruction manual before installing sensor in the tank.
- 4.2.5 Place the cell in the liquid and secure to the tank in an appropriate manner using the wire bracket provided with MHMSC hardware.

4.3 Electrical Connection

The sensor is electrically connected directly to the instrument or indirectly with a junction box and interconnect cable. A suitable junction box with a terminal strip is available from Water Analytics (Model JB-1).

4.3.1 Direct Hook-Up

- A. Route sensor cable to instrument.
- B. Connect sensor cable lead wires in accordance with instrument hook-up diagram.

Note: If you are using this conductivity cell with an AquaMetrix 2200C or 2100C instrument, connect the four wires to the cell terminals, matching the colors.

If you are using it with an instrument, which accepts 3 wires, connect as follows:

- Join the black wire and green wire together and connect to the black terminal.
- Connect the red and the white as usual to their respective terminals. CAUTION: If connecting to a GLI instrument, reverse the red and white wires.
- The shield is not required so there will be no connection to the SHEILD terminal.

4.3.2 Indirect Hook-Up with Junction Box

- A. Mount junction box with terminal strip (AquaMetrix Model JB-1) on flat surface such that its cover is removable when installed.
- B. Route sensor cable to junction box. Keep terminal strip dry to prevent problems caused by wet and/or corroded terminals.
- C. Route interconnect cable to junction box and instrument. It is recommended that this cable be run in 1/2" metal conduit for protection against moisture and mechanical damage.

Do not run line power in the same conduit with interconnect cable because "electrical noise" may interfere with the sensor signal.

- D. Connect sensor and interconnect cable lead wires by matching colors to the junction box terminal strip. Fasten cover onto junction box.
- E. Connect interconnect cable lead wires to instrument in accordance with instrument hook-up diagram.

4.3.2 Wire Designations

White / Black: Electrodes

Red / Green: Thermistor (Temperature compensation)

5.0 CLEANING THE SENSOR

- 5.1** Wipe the electrodes end with a soft cloth and rinse with clean water.
- 5.2** Clean the sensor electrodes in a soap solution using a soft cloth, brush, Q-tip or pipe cleaner. If soap solution cleaning is unsuccessful, place the sensor electrode in 10% HCl (hydrochloric acid) for 1 to 5 minutes to remove the contaminants that are acid soluble. Acid clean only if absolutely necessary. Never use abrasive cleaners on the sensor or its electrodes.

6.0 TROUBLESHOOTING AND SERVICE


- 6.1** Use the troubleshooting section in the analyzer instruction manual to determine whether the sensor or analyzer is defective.
- 6.2** Should service, parts, assistance in troubleshooting or repairs be required, please contact your Water Analytics representative or the factory directly:

Water Analytics Inc.
100 School Street
Andover, MA 01810

978-749-9949 Toll free: 855-747-7623
Fax: (905) 749-9961
Email: support@WaterAnalytics.com

An RMA# is required for all returning items. Please contact Water Analytics if you wish to send a product for repair or evaluation. A description of the malfunction as well as the proper return address should accompany all sensors returning for repair.

7.2 N105-60 INSTALLATION OF AS SERIES CONTACTING CONDUCTIVITY CELLS



REVISIONS			
SN	DESCRIPTION	DATE	APPROVAL
1	ORIGINAL ISSUE	3/14/03	MJCLAW

DATE	BY

AquaMetrix Inc.
Industrial and Environmental Controls

INSTALLATION OF MS SERIES CONTACTING CONDUCTIVITY CELLS

CODE DESIG. NO. 1 SDR
 QUANTITATIVE DRAWING NO. N105-60
 SHEET 03 OF 01

7.3 N105-61 ASSEMBLY OF AS SERIES CONTACTING CONDUCTIVITY CELLS COMPRESSION FITTING

